




Flower Plate

For Mediocarpal Partial Arthrodeses





Our core competence is hand surgery, a field where we can offer you much more than just standard treatment solutions for, say, distal radius fractures. Many of our products are intended to help you to achieve outstanding results in difficult, non-everyday situations as well. Products such as our ulnar head prosthesis (UHP) or the Flower Plate for mediocarpal partial arthrodesis are excellent examples of this.

Our objective is to simplify hand surgery interventions through intelligent system solutions, helping you to achieve the best possible results in the interest of the patient. Working in close cooperation with well-known authors and their teams, we have translated new ideas into innovative products that are consistently being developed further in an ongoing process. The result is a wide range of high-quality systems that impress with their clever design along with easy and safe handling.

And what's more, we have never lost sight of the economic perspective and service needs of our customers.

We consider ourselves as a true partner – to be relied upon for routine tasks and special challenges alike.

Table of Contents

	Pages
Product Features	6-9
Indications and Surgical Technique	10-21
■ Four Corner Fusion Treatment with the Flower Plate	12-21
Product Range	
■ Implants	22-23
■ Instruments	24-25
■ Storage System	26-27





Flower Plate

For Mediocarpal Partial Arthrodeses

Mediocarpal partial arthrodeses with Kirschner wires and the resulting plaster cast immobilization of the hand for a period of several weeks represent a considerable limitation for active patients.

Increased expectancy on the part of affected patients calls for the possibility of early mobilization with rapid restoration of normal function so patients are rehabilitated and can resume everyday life and work as soon as possible.

The Flower Plate, an alternative method of performing medio-carpal partial arthrodeses, satisfies this expectancy and meets the requirements in every respect.

Due to its small size and delicate shape, which is ideally matched to the carpal bones, it enables optimal placement below bone level in order to avoid impingement at the dorsal edge of the radial bone. The free combinability of standard screws and multidirectional locking screws ensures optimal fixation and hence a stable connection with the bones – an ideal scenario for early mobilization and rapid rehabilitation.

Feature, Function
and Benefit



The Flower Plate is impressive particularly because of its small size and delicate shape, which is ideally matched to the carpal bones. The two properties enable not only optimal placement of the plate but also positioning below bone level and minimization of the amount of bone to be removed – an ideal scenario to avoid painful impingement injuries at the dorsal edge of the radial bone.

Free use of standard screws and multidirectional locking screws in each plate hole also offers optimal fixation options to ensure a stable connection with the bones and hence promising prospects of regeneration.

Flower Plate



Scale 1:1



Scale 1:1


Features and functions

- Small plate size and delicate, concave shape

- Low plate profile


Benefits

- Prevents impingement at the dorsal edge of the radial bone
- Can be positioned below bone level
- Reduces the amount of bone to be removed




- Anatomical pre-shaped plate design
- Rounded, atraumatic plate contour
- Integrated center hole

- Ideally matched to the carpal bones, thus enabling accurate positioning
- Optimal embedding in soft tissue with maximum preservation
- Enables the introduction of additional cancellous bone



- Multidirectional locking plate holes


- Free combinability of standard screws and multidirectional locking screws in diameters 2.0 mm and 2.3 mm
- High intraoperative flexibility
- Optimal fixation with +/- 20° degree of angulation
- Stable connection with the bones
- Enables early mobilization and rapid rehabilitation



- Practical sterile kit including positioning instrument

- Enables easy positioning and orientation of the plate
- Simplifies placement of the first two screws in lunate and capitate (FCF fusion)
- Enables drilling, determination of screw length, and introduction of the screw via integrated guides

Feature, Function and Benefit



The Flower Plate can be fixated both with standard screws and with multidirectional locking screws in diameters 2.0 mm and 2.3 mm. The free combinability ensures optimal management and a stable connection with the bones so it provides an ideal scenario for early mobilization, rapid rehabilitation, and a painfree everyday life with maximum freedom of movement.

Clear identification of diameters is ensured with color-coded single clips.

Color code:	Screw diameter
Red:	2.0 mm
Black:	2.3 mm


Color-coded screws mean instant identification between standard and multidirectional locking screws, even when stored in the clip.

Color code	Screw	Diameter
Gold:	Standard screw	2.0 mm / 2.3 mm
Blue:	Locking screw	2.0 mm / 2.3 mm

smartDrive® Screws

Features and functions

Benefits




- Screw head and tip of atraumatic design

- Secure and soft-tissue-friendly bicortical anchorage in the bone

- Double, self-tapping thread


- Reduces screwing-in time by 50% and keeps the required effort to a minimum



- Multidirectional locking screws in diameters 2.0 mm and 2.3 mm

- Secure, multidirectional locking of the screw in the plate (+/- 20°)

- Maximum deflection without soft tissue irritation




- T6 with self-retaining function

- Easy pick-up, insertion, tightening or removal of the screw

- Direct force transfer from the screwdriver blade to the screw

- Optimal synergy of handling and force transfer



- Color-coded single clip

- Clear assignment of the appropriate screw diameter

- Direct, swift and application-oriented access

- 100% batch traceability

- Chargeable individually

- Easy recording of all implant data

Step by Step to Optimal Fixation


Indications

The Flower Plate is used for the treatment of degenerative or posttraumatic carpal arthroses that has arisen due to instability or carpal collapse following scaphoid pseudarthrosis (SNAC wrist) or following rupture of the scapholunate ligament (SLAC wrist).

It can also be used following failed partial arthrodeses (using K-wires) and for complex fractures in the intercarpal region.




FCF fusion



CMC fusion



STT fusion



SC fusion



SLC fusion



Surgical Technique


Four Corner Fusion

Treatment with the Flower Plate

Prof. H. Krimmer


Pages 12-21





Preoperative planning

Standard exposures are made of the wrist in the neutral position with an A/P and lateral beam.




Patient positioning

The patient is placed in the supine position on the operating table. The hand to be operated on is placed in the pronation position of the lower arm on the hand side table, with complete deprivation of blood in the upper arm. The wrist is supported with a rolled-up towel.

Note:

The technique involves removal of the scaphoid and arthrodesis of lunate, capitate, hamate, and triquetrum. With this operation special attention must be paid to exact reduction of capitate and lunate in order to ensure proper wrist alignment and bone consolidation.



1. Approach

Opening is performed by making a slightly curved incision, about 6 cm long, above the dorsal wrist.




2. Exposure of the extensor tendon retinaculum

After skin incision the extensor tendon retinaculum is exposed.



3. *Opening of the extensor tendon compartments*

After release of the retinaculum the third extensor tendon compartment is opened toward radial and ulnar. The second and fourth extensor tendon compartment are also exposed.




4. *Opening of the wrist capsule*

The second extensor tendon compartment is held aside toward radial and the fourth extensor tendon compartment is held aside toward ulnar.

The dorsal interosseous nerve is exposed and extensively resected.

The wrist capsule is opened transversely in the direction of the dorsal ligaments, preferably through the pedunculated capsular ligamentous flaps according to Bishop and Berger.






5. *Excision of the scaphoid*

The scaphoid is exposed and removed in its entirety. In doing so, injury of the palmar ligaments must be prevented.

Removal of the scaphoid can optionally be performed with the aid of the CarpalStick, which, like a corkscrew, grips deeply into the structure of the bone. In this way high forces can be transferred to the threads of the instrument via the T-handle during removal.



6. *Cartilage removal*

The four bones to be fused, lunate, capitate, hamate, and triquetrum, are exposed. To enable them to grow together optimally, cartilage is removed from the individual articular surfaces using a Luer rongeur until the cancellous structures are reached.



CarpalStick




7. Preparation of the bone bed

When all the cartilage has been removed, the bony structures are freshened slightly. For this purpose a 1.5 mm diameter core hole drill can be used.

The base of the four bones to be fused is filled up with cancellous bone removed from the distal radius.



Alternatively, it is also possible to use cancellous bone harvested from the scaphoid or iliac crest.




8. Reduction of rotational deformities

Existing deformities, such as instability-related rotation of the individual carpal bones, are reduced. Alignment of the lunate is particularly important.

If necessary, the bones can be aligned by introducing a K-wire according to the joystick method.






9. Temporary fixation

Under x-ray control, the radius, lunate, and capitate are fixated temporarily using a Kirschner wire. Care must be taken to ensure that the Kirschner wire is placed as palmar as possible in order to avoid hindrances when reaming.

Alternatively, capitate and lunate can be fixated percutaneously from distal.

Note:

Temporary fixation reduces displacement when reaming. In addition, the resistance for the reamer is higher so a uniform, circular implant bed can be created.



10. Reaming of the implant bed



The self-centering reamer is placed over the four carpal bones to be fused, lunate, capitate, hamate, and triquetral. Ideally, the position is selected slightly distal in order to avoid impingement at the dorsal edge of the radial bone during extension of the wrist.

If necessary, the 1.5 mm diameter core hole drill can be used to pre-drill a minimum grit size for the tip of the reamer (2 mm max.) in order to facilitate correct placement of the reamer.

The bone surface is prepared with the reamer until the edge of the reamer is flush with the dorsal bone surface. The Flower Plate is inserted for a trial in order to check the depth of the implant bed. This ensures that the plate can be positioned below bone level.



Reamer



11. Placement of the Flower Plate

The Flower Plate is introduced using the positioning instrument and aligned in such a way that one screw can be placed in the lunate and one in the capitate.

Note:

For optimal screw placement there are two screws in the lunate, one screw in the triquetral, two screws in the capitate, and one screw in the hamate.

The standard length is usually 10 - 16 mm.

12. Drilling the first core hole

For the drilling procedure the drill sleeve is first inserted in the drill guide of the positioning instrument leading to the lunate. The core hole drill is introduced via the opening of the drill sleeve and placed on the bone.

That is followed by monocortical pre-drilling. After drilling, the drill sleeve is removed.




Flower Plate sterile kit



Flower Plate




Core hole drill, Ø 1.5 mm



13. Determination of screw length

Correct screw length is determined with the depth gauge, which is used for screw diameters 2.0 mm and 2.3 mm. The depth gauge is introduced via the opening in the drill guide.

The screw selected must be 2 mm shorter than the value measured with the depth gauge.



14. Placement of the first screw


The plate is first fixated in the lunate with a standard screw. For this purpose the screw is picked up with the color-coded screwdriver, which is used for the two diameters 2.0 mm and 2.3 mm, and driven into the bone via the opening in the drill guide.



Depth gauge 2.0 mm, 2.3 mm




T6 screwdriver



15. Placement of the second screw

Successful implantation of the first screw in the lunate is followed by fixation of the plate in the capitate. For this purpose the drill sleeve is first inserted in the drill guide leading to the capitate. The second standard screw is then placed by means of the technique described in steps 12-14. At this point it is advisable to subject the plate to a radiological and functional check. Following a successful check and secure fixation of the two first screws the positioning instrument is removed.




16. Placement of further screws


To achieve adequate stability more plate holes are filled with screws. Optionally, multidirectional locking screws can be used to increase stability. The procedure for this is described in steps 12-14, although the drilling is performed without the drill guide or drill sleeve.

If necessary, the two standard screws can also be replaced by multidirectional locking screws.

As required, more cancellous bone can be introduced via the circular opening at the center of the plate.




T6 screwdriver



17. Wound closure

After irrigation and cleaning of the wound the capsular and ligamentous structures are carefully restored.

That is followed by skin suture.



18. After-treatment

After skin closure a final x-ray image is made.

Following surgery, the patient should wear, for protection, a forearm splint that allows active finger movement, for about 4 weeks. Physical therapy can commence 2 weeks after surgery, wearing the splint.

With regular x-ray exposures the progress of desired fusion is checked prior to the resumption of normal activities.

Flower Plate Implants

Flower Plate


7-hole

Profile height 3.7 mm

Flower Plate sterile kit

consisting of:

- Flower Plate
- Positioning instrument
- Drill sleeve



1/1

26-130-07-09

26-130-07-71

STERILE | R



= 1.2 mm



1/2

26-130-17-71

STERILE | R



Explanation of icons

-  Screw diameter 2.0 mm
-  Screw diameter 2.3 mm
-  Titanium
-  T-Drive
-  Packaging unit

STERILE | R Sterile packaged implants

Ø 2.0 mm

Standard screw	Multidirectional locking screw
	
$\frac{1}{1}$	$\frac{1}{1}$

Length	Item No.	STERILE R	Item No.	STERILE R
8 mm	26-020-08-91	26-020-08-71	26-019-08-91	26-019-08-71
10 mm	26-020-10-91	26-020-10-71	26-019-10-91	26-019-10-71
12 mm	26-020-12-91	26-020-12-71	26-019-12-91	26-019-12-71
14 mm	26-020-14-91	26-020-14-71	26-019-14-91	26-019-14-71
16 mm	26-020-16-91	26-020-16-71	26-019-16-91	26-019-16-71
18 mm	26-020-18-91	26-020-18-71	26-019-18-91	26-019-18-71
20 mm	26-020-20-91	26-020-20-71	26-019-20-91	26-019-20-71

Ø 2.3 mm

Standard screw	Multidirectional locking screw
	
$\frac{1}{1}$	$\frac{1}{1}$

Length	Item No.	STERILE R	Item No.	STERILE R
8 mm	26-023-08-91	26-023-08-71	26-022-08-91	26-022-08-71
10 mm	26-023-10-91	26-023-10-71	26-022-10-91	26-022-10-71
12 mm	26-023-12-91	26-023-12-71	26-022-12-91	26-022-12-71
14 mm	26-023-14-91	26-023-14-71	26-022-14-91	26-022-14-71
16 mm	26-023-16-91	26-023-16-71	26-022-16-91	26-022-16-71
18 mm	26-023-18-91	26-023-18-71	26-022-18-91	26-022-18-71
20 mm	26-023-20-91	26-023-20-71	26-022-20-91	26-022-20-71

Note: The screw modules can each store a total of 60 screw clips in screw lengths 8, 10, 12, 14, 16, 18, 20 mm, with 4 standard screws and 4 locking screws per indicated length.

You will find all the intermediate lengths in the brochure "LINOS Hand Fracture System".

Flower Plate Instruments

Standard instrumentation



1/2

26-130-18-07
K-wire dispenser
Ø 1.6 mm
18 cm



1/2

22-626-16-05
K-wires
Ø 1.6 mm
12 cm



1/2

26-130-13-07
Reamer
Ø 13 mm
AO attachment



1/2

26-153-16-07
26-153-16-71
Core hole drill
Ø 1.5 mm



STERILE R



1/2

26-153-18-07
26-153-18-71
Core hole drill
Ø 1.8 mm



STERILE R



Explanation of icons

- Screw diameter 2.0 mm
- Screw diameter 2.3 mm
- Steel
- Silicone
- T-Drive
- Dental attachment
- AO attachment
- Packaging unit

STERILE Sterile packaged instruments

STERILE Steam-sterilized

Optional instruments



1/2

26-975-30-07
Depth gauge
Ø 2.0/2.3 mm
One-handed design



1/2

26-975-36-07
Screwdriver
T6
Short, rotatable



1/1

26-975-02-04
Screw measurement clip
Length and diameter



1/2

23-192-00-71
CarpalStick
8 cm



STERILE

Flower Plate Storage

The storage system is impressive not only because of its easy, well-conceived handling, for example with the instruments arranged according to their sequence of use during surgery, but also because of its optimized reprocessing capability, in order to equally serve the needs of everyone involved.

In the storage basket all the instruments required for an operation can be stored individually next to each other. It is also possible to accommodate multiple Flower Plate implants.

Every screw module can accommodate a total of 60 screws in lengths ranging from 8 to 20 mm, all stored in single clips. The clips, which are labeled with screw length and diameter, article number, and batch number, permit not only easy recording of all the relevant implant data but also seamless patient-related documentation.

Apart from the option of conventional storage, the Flower Plate system is also available with sterile packaged implants throughout.



55-910-81-04 Instrument storage complete, comprised of:

55-910-89-04 Storage cage

55-910-59-04 Lid



55-910-89-04
Storage cage



55-910-59-04
Lid

Screw module combination*

55-910-82-04 Screw module, combination of standard screws/locking screws, Ø 2.0 mm

55-910-83-04 Screw module, combination of standard screws/locking screws, Ø 2.3 mm



55-910-82-04
Screw module, standard screws/locking screws, Ø 2.0 mm



55-910-83-04
Screw module, standard screws/locking screws, Ø 2.3 mm

* **Note:** The screw modules can each store a total of 60 screw clips in screw lengths 8, 10, 12, 14, 16, 18, 20 mm, with 4 standard screws and 4 locking screws per indicated length.

KLS Martin Group

KLS Martin France SARL
68200 Mulhouse · France
Tel. +33 3 89 51 31 50
france@klsmartin.com

KLS Martin UK Ltd.
Reading RG1 3EU · United Kingdom
Tel. +44 1189 000 570
uk@klsmartin.com

KLS Martin do Brasil Ltda.
CEP 04.531-011 São Paulo · Brazil
Tel.: +55 11 3554 2299
brazil@klsmartin.com

Gebrüder Martin GmbH & Co. KG
Representative Office
121471 Moscow · Russia
Tel. +7 499 792-76-19
russia@klsmartin.com

Martin Italia S.r.l.
20864 Agrate Brianza (MB) · Italy
Tel. +39 039 605 67 31
italia@klsmartin.com

Nippon Martin K.K.
Tokyo 113-0024 · Japan
Tel. +81 3 3814 1431
nippon@klsmartin.com

KLS Martin Australia Pty Limited
Artarmon NSW 2064 · Australia
Tel.: +61 2 9439 5316
australia@klsmartin.com

Gebrüder Martin GmbH & Co. KG
Representative Office
201203 Shanghai · China
Tel. +86 21 5820 6251
china@klsmartin.com

Martin Nederland/Marned B.V.
1271 AG Huizen · The Netherlands
Tel. +31 35 523 45 38
nederland@klsmartin.com

KLS Martin L.P.
Jacksonville, FL 32246 · USA
Tel. +1 904 641 77 46
usa@klsmartin.com

KLS Martin SE Asia Sdn. Bhd.
14100 Simpang Ampat, Penang · Malaysia
Tel.: +604 505 7838
malaysia@klsmartin.com

Gebrüder Martin GmbH & Co. KG
Representative Office
Dubai · United Arab Emirates
Tel. +971 4 454 16 55
middleeast@klsmartin.com

Gebrüder Martin GmbH & Co. KG
A company of the KLS Martin Group
KLS Martin Platz 1 · 78532 Tuttlingen · Germany
Postfach 60 · 78501 Tuttlingen · Germany
Tel. +49 7461 706-0 · Fax +49 7461 706-193
info@klsmartin.com · www.klsmartin.com

